

There are two simple checks which station staff should carry out during routine station patrols as follows:

1. Ensure that the green seal on the cabinet door is intact:



2. Check that a green tick (✓) is displayed on the defibrillator – this should be clearly visible without opening the cabinet door:



If either the seal is broken, or the green tick is not displayed, then please report this immediately to Customer Service Control and your Station Manager.